

**MANUFACTURING :**

Custom - built Plasma &amp; Vacuum Components and Assemblies

**SERVICES :**Helium Leak Testing, Methane Leak Testing as per API & ISO-15848 Standards  
Vacuum Brazing, Vacuum Heat Treatment, Calibration of Vacuum Gauges

Representative of Saes Getters, Italy for Vacuum Pumps in India

Plot No. 17, Road 1 - A,  
GIDC, Kathwada, Ahmedabad-382430.  
Phone : +91 79-29705319, Fax : 079-29709216  
Email : [pvt@plasmavac.com](mailto:pvt@plasmavac.com)  
Website : [www.plasmavac.com](http://www.plasmavac.com)

Certificate No : 624202202275

Date : February 27, 2022

**API STANDARD 624 TEST CERTIFICATE****Type Testing of Rising Stem Valves Equipped with Graphite Packing For Fugitive Emissions**

This is to certify that the below mentioned valve has been inspected by Plasma &amp; Vacuum Technologies and found to be satisfactory in accordance with the requirements of API 624, First Edition 2014.

Manufacturer:	<b>Fluid-O-Mech Controls Inc.</b>
Address:	<b>14/1A, Pancharatna Industrial Estate, Near Ramol Bridge, Vatva G.I.D.C, Ahmedabad – 382 445. Gujarat. India.</b>
Location of Test:	<b>Plasma &amp; Vacuum Technologies, Plot No.17, Road 1-A, GIDC Kathwada, Ahmedabad 382430, India.</b>

Product :	Valve Details:	<b>Globe Valve, 1 1/2" 800#</b>
	Design Standard:	<b>API 602</b>
Packing :	<b>Make: New Empire Gaskets, Model: Seal-It.</b>	
Body/Bonnet Material :	<b>ASTM A105</b>	
Stem Material :	<b>ASTM A276 TP410</b>	
GA Drawing No:	<b>FMC/40X800#/GBV/SW/00</b>	
Test Fluid :	<b>Methane (&gt;97 % Purity)</b>	
Test Pressure :	<b>41.4 barG (~600 psiG)</b>	
Test Temperature :	<b>At Ambient and 260°C (~ 500°F)</b>	
Mechanical Cycle :	<b>310 Cycles over 3 Thermal Cycles</b>	
Maximum Leakage :	<b>59.7</b>	<b>ppmv</b>
Average Leakage :	<b>31.9</b>	<b>ppmv</b>
Acceptance Criteria :	<b>&lt; 100</b>	<b>ppmv maximum</b>
Test Result :	<b>PASS</b>	
Test Report No :	<b>PVT/SD/FET/R/202202/26</b>	
Test Start Date :	<b>February 23, 2022</b>	
Test End Date :	<b>February 25, 2022</b>	

Valve Qualified : All valves of the same basic design as the test valve may be deemed to have been type tested, subject to the following additional limitation. For API 602 Valves, 1 1/2" (DN 40) Class 800 test valve qualifies all valves from 1 1/4" (DN 32) through 2 1/2" (DN 65) and up to and including pressure class 800.

Test Conducted by	Test Witnessed By
Venkat N. Ramani ASNT Level III (LT) 183918	Witnessed 27/02/2022





## MANUFACTURING :

Custom - built Plasma & Vacuum Components and Assemblies

## SERVICES :

Helium Leak Testing, Methane Leak Testing as per API & ISO-15848 Standards

Vacuum Brazing, Vacuum Heat Treatment, Calibration of Vacuum Gauges

Plot No. 17, Road 1 - A,

GIDC, Kathwada, Ahmedabad-382430.

Phone : +91 79-29705319, Fax : 079-29709216

Email : [pvt@plasmavac.com](mailto:pvt@plasmavac.com)

Website : [www.plasmavac.com](http://www.plasmavac.com)

Saes  
getters

Representative of Saes Getters, Italy for Vacuum Pumps in India

## API STANDARD 624 FUGITIVE EMISSIONS TEST REPORT

F/CR/10.2  
RevNo.00

Report No :	PVT/SD/FET/R/202202/26 ✓	Report Date :	27 February 2022 ✓
Manufacturer :	Fluid-O-Mech Controls Inc.	Valve Stem Type :	Rising & Rotating Stem ✓
Valve - Type, Size, Class :	Globe Valve, 1 1/2" 800# ✓	API / ASME Design Standard:	API 602
Packing Description :	Number of Rings : 6 nos OD : 25.0 mm, ID : 15.25 mm Thickness : 4.5 mm Stack Height : 27 mm Stuffing box depth : 28 mm		
Make: New Empire Gaskets, Model: Seal-It.	Body / Bonnet Gasket Connection: SPW SS316 + Graphite Filler.		
Packing Manufacturer :			
New Empire Gaskets ✓	Gland Follower Exposed Length:17mm	Body-Bonnet connection : Bolted.	
Gland Nut recommended torque:38Nm	Gland Follower Insertion Depth: 2mm	Body/Bonnet Nut recommended torque: 160Nm	
Test Start Date :	23 February 2022	Valve Selection : by Manufacturer; Selection Date:17.02.2022	
Test Completion Date :	25 February 2022	Selected by : QA/QC Dept. Fluid-O-Mech Controls Inc. Ahmedabad.	

### SUMMARY OF FUGITIVE EMISSIONS TEST DATA

Test Segment	Cycle	Methane Pressure (barG)	Temperature at Body (°C)	Temperature at Stem (°C)	Static Leak Measurement (ppmv)	Dynamic Leak Measurement (ppmv)	Remarks
Ambient Temperature T(a) = RT, P(a)=41.4 barG	0	41.4	28	28 ✓	13.4	24.2	OK
	50	41.4	29	30	18.7		
Elevated Temperature T(e) =260°C, P(e) = 41.4 barG	51	41.4	262	258 ✓	20.9	29.0	OK
	100	41.4	264	261	25.2		
Ambient Temperature T(a) = RT, P(a)=41.4 barG	101	41.4	29	30	23.5	32.9	OK
	150	41.4	31	32	28.2		
Elevated Temperature T(e) =260°C, P(e) = 41.4 barG	151	41.4	260	259	25.3	49.1	OK
	200	41.4	265	263	45.2		
Ambient Temperature T(a) = RT, P(a)=41.4 barG	201	41.4	27	27	24.6	41.6	OK
	250	41.4	29	29	35.0		
Elevated Temperature T(e) =260°C, P(e) = 41.4 barG	251	41.4	267	268	28.6	59.7	OK
	300	41.4	269	269	56.1		
Ambient Temperature T(a) = RT, P(a)=41.4 barG	301	41.4	39	40	22.4	34.3	OK
	310	41.4	39	40	32.7		

Valve Serial Number: 3H442 (Heat No.: Body: E1221, Bonnet: B242) ✓

Running Torque - First Cycle : 12 Nm, Last cycle : 10 Nm

FINAL REMARK : Pass (✓)

Pre-test Preparations & Adjustments : 1) Gland bolting, Body/Bonnet and Valve open-close torques were verified; 2) The air in the valve cavity was evacuated and purged with Helium prior to starting the testing; 3) The stem orientation was kept vertical; 4) External valve heating was designed, with heating coils and wrapped by Alumina wool; 5) Electrical Operation using wheel was set for cycling (opening and closing) the valves.

Notes : 1) The test valve was randomly selected; 2) The test medium was methane with purity > 97%; 3) The test was conducted in a safe, well ventilated and protected environment; 4) No packing adjustment was done during test. Running torques at the start and end were measured and recorded; 5) Methane Detector (Model TVA 2020, Thermofisher) was used for ambient monitoring; 6) Methane Detector (Model TVA 2020, Thermofisher) was used for methane leakage at gland and body / bonnet joint. The Probe was calibrated prior to each measurement using an external calibration gas with known Methane concentration. T(a)= Ambient Temperature, T(e)= Elevated Temperature, P(a)= Ambient Pressure & P(e)= Elevated Pressure, RT=Room Temperature

Test Witnessed by

Test Conducted by

Test Witnessed By

Venkat N. Ramani  
ASNT Level III (LT)  
Plasma & Vacuum Technologies







## Annexure 1 : SET UP

Report No.:

PVT/SD/FET/R/202202/ 26

Report Date: 27 February 2022



(W)



Shakil S

1 : Methane Gas Cylinder; 2 : Isolation Valve; 3 : Pressure Gauge;

4 : Valve under Test; 5 : Temperature Sensor; 6 : Hot Zone;

7 : Methane Leak Detector; 8 : Sniffer Probe; 9 : Valve operation-Rising & Rotating Stem.




## Annexure 2

### Schematic of Probe Calibration Set up

Report No.:

PVT/SD/FET/R/202202/ 26

Report Date: 27 February 2022



1 : Methane Gas Cylinder; 2 : Zero Gas / Helium Cylinder; 3,4,5 : Positions of Methane Leak Detector. For Calibration, EPA21 Calibration and Zeroing respectively; 6 : Vacuum Pump; 7 : Reservoir; 8, 9, 10, 11 : Pressure Gauges; 12 : Methane Calibrated Leak.



W







## ANNEXURE 3 : TEST REPORT - HELIUM LEAK TESTING

Report No :	PVT/SD/FET- HLT/R/202202-26 ✓	Report Date :	February 23, 2022 ✓
Client :	Fluid-O-Mech Controls Inc.	Testing Equipment :	HLD ASM 310 ADIXEN
Manufacturer :	Fluid-O-Mech Controls Inc.	Calibrated Leak(s) :	PVT/CD/SL/03, PVT/CD/SL/04
Item :	Globe Valve, 1 1/2" 800#	Ref. Code(s) :	ASME Sec V, Art 10, Detector Probe Technique
Reference:	Prior to API624 Test of valve	Qualification:	Helium Leakage Rate : <1.0x10 <sup>-6</sup> mbar l/s at Body-Bonnet and Gland joints at 41.4 barG pressure (10% of Helium + 90% Nitrogen).

### Leak Detector and Probe Calibration

	Standard Leak Value		Observed Leak Value		Instrument Sensitivity		
A	$2.48 \times 10^{-8}$	mbar l/s	$\sim 2.5 \times 10^{-8}$	mbar l/s	better than	$10^{-9}$	std cc/s
B	$1.13 \times 10^{-6}$	mbar l/s	$\sim 1.1 \times 10^{-6}$	mbar l/s	CF = 1.0	Response	< 5 s
Stage	Test Date	Body		Gland Joint		Remarks	
		Leak Rate (mbar l/s)		Leak Rate (mbar l/s)			
1	Prior to API624 Test	Feb 23, 2022	$8.7 \times 10^{-7}$	$9.6 \times 10^{-7}$	Leak Tightness is OK		

Test Witnessed by	Test Conducted by	Test Witnessed By
Fluid-O-Mech Controls Inc.	Venkat N. Ramani ASNT Level III (LT) Plasma&Vacuum Technologies	  23/02/2022 Shaikil S.



## MANUFACTURING :

Custom - built Plasma & Vacuum Components and Assemblies

## SERVICES :

Helium Leak Testing, Methane Leak Testing as per API & ISO-15848 Standards

Vacuum Brazing, Vacuum Heat Treatment, Calibration of Vacuum Gauges

Plot No. 17, Road 1 - A,

GIDC, Kathwada, Ahmedabad-382430.

Phone : +91 79-29705319, Fax : 079-29709216

Email : [pvt@plasmavac.com](mailto:pvt@plasmavac.com)

Website : [www.plasmavac.com](http://www.plasmavac.com)

**Saes** getters Representative of Saes Getters, Italy for Vacuum Pumps in India

## Annexure 4

## CALIBRATION DATA AND METHANE LEAKAGE EVALUATIONS

Report No.: **PVT/SD/FET/R/202202/ 26**

Report Date: **27 February 2022**

Test Segment	Cycle	Condition	Ambient Leak Rate Value (ppmv)	Calibration Leak Rate Value (ppmv)	Detector Sensitivity	Observed Leak Rate (ppmv)	Measured Leak Rate (ppmv)
Ambient Temperature T(a) = RT, P(a)=41.4 barG	0	Static	2.3	75.4	1.02	15.4	13.4
	50	Static	4.2		1.05	22.1	18.7
	50	Dynamic				27.3	24.2
Elevated Temperature T(e) =260°C, P(e) = 41.4barG	51	Static	5.1	75.8	1.05	24.9	20.9
	100	Static	7.6		1.09	30.7	25.2
	100	Dynamic				34.1	29.0
Ambient Temperature T(a) = RT, P(a)=41.4 barG	101	Static	2.8	76.3	1.01	26.0	23.5
	150	Static	5.2		1.05	32.1	28.2
	150	Dynamic				36.6	32.9
Elevated Temperature T(e) =260°C, P(e) = 41.4barG	151	Static	8.0	76.7	1.08	31.3	25.3
	200	Static	3.4		1.02	47.9	45.2
	200	Dynamic				51.7	49.1
Ambient Temperature T(a) = RT, P(a)=41.4 barG	201	Static	7.0	74.9	1.10	29.4	24.6
	250	Static	2.1		1.02	36.3	35.0
	250	Dynamic				42.7	41.6
Elevated Temperature T(e) =260°C, P(e) = 41.4barG	251	Static	7.7	75.4	1.10	33.7	28.6
	300	Static	5.9		1.07	58.2	56.1
	300	Dynamic				61.6	59.7
Ambient Temperature T(a) = RT, P(a)=41.4 barG	301	Static	8.4	75.9	1.10	28.7	22.4
	310	Static	2.8		1.02	34.9	32.7
	310	Dynamic				36.4	34.3
Average Value:						31.9	

Calibrated Leak used (ml/m) : SL 0.072

Flow Rate at Methane Leak Detector (lpm) : FR 0.966

Estimated PPM value (PPMv) = (SL/FR)x1000 74.53

## Leakage at Body-Bonnet Connection

Cycle Number	Body Temperature (°C)	Pressure (barG)	Ambient Leak Rate Value (ppmv)	Calibration Leak Rate Value (ppmv)	Detector Sensitivity	Observed Leak Rate Value (ppmv)	Measured Leak Rate Value (ppmv)
0	28	41.4	2.3	75.4	1.02	18.9	16.9
310	39	41.4	2.8	75.9	1.02	41.3	39.3

Measured Leak Rate = (Observed Leak Rate – Ambient Leak Rate Value) x  
[Estimated PPM value / (Calibration Leak Rate Value – Ambient Leak Rate Value)]

Test Witnessed by


Fluid-O-Mech Controls Inc.

Test Conducted by

Venkat N. Ramani  
ASNT Level III (LT)  
Plasma&Vacuum Technologies



Test Witnessed By







## Annexure 5 Methane Leakage Measurements

Report No. : PVT/SD/FET/R/202202/ 26


Report Date: 27 February 2022




(a)



(b)




(c)



(d)

Methane leakage at : a) Ambient, Dynamic, 50 cycles; b) Elevated, Dynamic, 100 cycles; c) Elevated, Dynamic, 200 Cycles ;  
d) Elevated, Dynamic 300 Cycles ;





## MANUFACTURING :

Custom - built Plasma & Vacuum Components and Assemblies

## SERVICES :

Helium Leak Testing, Methane Leak Testing as per API & ISO-15848 Standards  
Vacuum Brazing, Vacuum Heat Treatment, Calibration of Vacuum Gauges

Plot No. 17, Road 1 - A,

GIDC, Kathwada, Ahmedabad-382430.

Phone : +91 79-29705319, Fax : 079-29709216

Email : [pvt@plasmavac.com](mailto:pvt@plasmavac.com)

Website : [www.plasmavac.com](http://www.plasmavac.com)

Saes  
getters



Representative of Saes Getters, Italy for Vacuum Pumps in India

## Annexure 6


### Temperature Measurements

Report No. : PVT/SD/FET/R/202202/ 26

Report Date: 27 February 2022



1 1/2 Inch 800# Globe Valve ; Fluid-O-Mech Controls Inc.  
API 624 Test : Temperature Profile



(W)





**MANUFACTURING :**  
Custom - built Plasma & Vacuum Components and Assemblies

**SERVICES :**  
Helium Leak Testing, Methane Leak Testing as per API & ISO-15848 Standards  
Vacuum Brazing, Vacuum Heat Treatment, Calibration of Vacuum Gauges


Plot No. 17, Road 1 - A,  
GIDC, Kathwada, Ahmedabad-382430.  
Phone : +91 79-29705319, Fax : 079-29709216  
Email : [pvt@plasvac.com](mailto:pvt@plasvac.com)  
Website : [www.plasvac.com](http://www.plasvac.com)

**Saes**  
getters  
Representative of Saes Getters, Italy for Vacuum Pumps in India


## Annexure 7 Pressure Measurements

Report Date: PVT/SD/FET/R/202202/ 26


Report Date: 27 February 2022



1 1/2 Inch 800# Globe Valve ; Fluid-O-Mech Controls Inc.  
API 624 Test : Pressure Profile



W





## MANUFACTURING :

Custom - built Plasma & Vacuum Components and Assemblies

## SERVICES :

Helium Leak Testing, Methane Leak Testing as per API & ISO-15848 Standards  
Vacuum Brazing, Vacuum Heat Treatment, Calibration of Vacuum Gauges

Saes  
getters

Representative of Saes Getters, Italy for Vacuum Pumps in India

Plot No. 17, Road 1 - A,

GIDC, Kathwada, Ahmedabad-382430.

Phone : +91 79-29705319, Fax : 079-29709216


Email : [pvt@plasmavac.com](mailto:pvt@plasmavac.com)

Website : [www.plasmavac.com](http://www.plasmavac.com)



## Annexure 8 Valve Operation Mechanism

Report No.: PVT/SD/FET/R/202202/ 26

Report Date: 27 February 2022



API 624 Test : Methane Leakage Measurement Points



W





## MANUFACTURING :

Custom - built Plasma & Vacuum Components and Assemblies

## SERVICES :

Helium Leak Testing, Methane Leak Testing as per API & ISO-15848 Standards

Vacuum Brazing, Vacuum Heat Treatment, Calibration of Vacuum Gauges

Saes  
getters

Representative of Saes Getters, Italy for Vacuum Pumps in India

Plot No. 17, Road 1 - A,

GIDC, Kathwada, Ahmedabad-382430.

Phone : +91 79-29705319, Fax : 079-29709216

Email : [pvt@plasmavac.com](mailto:pvt@plasmavac.com)



Website : [www.plasmavac.com](http://www.plasmavac.com)

## Annexure 9

### Post – Test Valve Opened Up Photographs

Report No.: PVT/SD/FET/R/202202/ 26

Report Date: 27 February 2022



W

27/02/2022  
Shakti S



#### Remarks:

The disassembled Valve parts were inspected and the condition was found satisfactory.



**MANUFACTURING :**

Custom - built Plasma &amp; Vacuum Components and Assemblies

**SERVICES :**

Helium Leak Testing, Methane Leak Testing as per API &amp; ISO-15848 Standards

Vacuum Brazing, Vacuum Heat Treatment, Calibration of Vacuum Gauges

Plot No. 17, Road 1 - A,

GIDC, Kathwada, Ahmedabad-382430.

Phone : +91 79-29705319, Fax : 079-29709216

Email : [pvt@plasvac.com](mailto:pvt@plasvac.com)Website : [www.plasvac.com](http://www.plasvac.com)

Representative of Saes Getters, Italy for Vacuum Pumps in India

## List of Calibration Reports of Instruments / Documents attached

Report No.: PVT/SD/FET/R/202202/ 26

Report Date: 27 February 2022

- 1 Methane Gas Cylinder 709
- 2 Gas Monitor Pump (Flow Rate), Report No. PVT/SD/FRM/R/2022/01
- 3 Standard Leak, SI No PVT/TD/CL/06, Report No. PVT/SD/FRM/R/2022/02
- 4 Helium Standard Leak, SI No. PVT/CD/SL/03
- 5 Helium Standard Leak, SI No. PVT/CD/SL/04
- 6 Torque Wrench, MACMASTER, PVT/SD/TW/01 & 02
- 7 Pressure Gauge, LA, PVT/SD/GP70D/16 ✓
- 8 Temperature Sensor, Thermal Sense Tech, PVT/SD/TST/28 & 29 ✓
- 9 ASNT NDT Level III certification of V N Ramani
- 10 API 622 Qualified Packing Certificate

Page 12 of 12



Verified  
27/02/2022  
Shamir S







Since 1998

# NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagal Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058

Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370

E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com

Visit our Web Site : [www.calibrationlaboratory.in](http://www.calibrationlaboratory.in)



CC-2128

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

## Calibration Certificate

Name of Customer → <b>Plasma &amp; Vacuum Technologies</b> Plot No.17, 1/A Road, GIDC Kathwada, Ahmedabad-382 430, Gujarat.	Certificate No. NCQC-M/161221/03 Date of Issue 17-12-2021 Date of Calibration 16-12-2021 Suggested Due Date 15-12-2022
---	---

Date Of Receipt / Ref. No. → 16-12-2021

F/CR/M/052, Issue No.01  
Page 1 of 1

ULR – CC212821000005773F

Discipline → Mechanical Calibration,  
Force, Torque

Details of Observation of Unit Under Calibration

Identification No. : PVT/SD/TW/01  
Serial No. : 24B – 41  
Name of Instrument : Torque Wrench

Range	70 – 340 Nm	Visual Inspection	OK
Least Count	10 Nm	Make	Mac Master
Type	Type II	Model	TW 250 R
Accuracy	± 4 % of rdg.	Class	A

Force Set on Torque Wrench in Nm	Reading Observed By Torque Wrench calibration system in Nm	Absolute Error In Nm	Expanded Uncertainty (±)
110	107.9	2.1	2.42 % rdg.
230	226.7	3.3	2.42 % rdg.
320	324.3	4.3	2.42 % rdg.

### Remarks:

- Averages of minimum five readings are reported.
- Suggested due date is given based on customer requirements.
- Calibration points are given based on customer requirements.
- These results are obtained at the time of calibration.
- Any hand written corrections (except @ marked) or photocopies of the report invalidates this certificate.
- The uncertainties are for a confidence probability of not less than 95% with coverage factor k = 2.
- Environment condition during calibration: 23 ± 2°C, 40 to 60% Rh.
- Reference standard no.: IS/ISO 6789.
- Location of performance of calibration → At Lab.
- Condition of instrument found satisfactory during receipt.
- Reference calibration method no.: NCQC/CM/M/052.
- Our masters are directly calibrated through NABL accredited calibration laboratory having direct Traceability with national / international standard.

### Details of master Instrument Used for Calibration

Nomenclature	Make / Model	Id. No. / Sr. No.	Calibration Due Date
Torque Wrench calibration system	Sushma / TS-103F & TDU-RB-103	TS13-0089,0091,0092,0093, / DU13-0032 (NCQC-M/022)	30-03-2023

NCQC System Certificate No. Certificate no. & traceability of master with National Standards

152, 152/2, 152/3, 152/4

Our master torque wrench calibration system is calibrated and traceable to National Standard through NABL accredited laboratory Sushma Industries calibration centre, Certificate no. SCPL/CC/2876/2877-2878, 2879 /03/2020-2021, Date - 31-03-2021.

(R)

23/12/2022  
Shachi S.

Traceable To National / International Standards.

Calibrated By

J. Khatari  
Jaydeep Khatri

Reviewed & Approved By

Jigar Panchal

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"





Since 1998

# NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058

Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370

E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com

Visit our Web Site : www.calibrationlaboratory.in



CC-2128

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

## Calibration Certificate

Name of Customer → **Plasma & Vacuum Technologies**  
Plot No.17, 1/A Road, GIDC Kathwada,  
Ahmedabad-382 430, Gujarat.

Certificate No. NCQC-M/161221/01  
Date of Issue 17-12-2021  
Date of Calibration 16-12-2021  
Suggested Due Date 15-12-2022  
F/CR/M/052, Issue No.01  
Page 1 of 1

Date Of Receipt / Ref. No. → 16-12-2021

ULR – CC212821000005771F

Discipline → Mechanical Calibration,  
Force, Torque

Details of Observation of Unit Under Calibration

Identification No. : PVT/SD/TW/02  
Serial No. : 24R – 12  
Name of Instrument : Torque Wrench

Range 10 – 68 Nm  
Least Count 2 Nm  
Type II  
Accuracy ± 4 % of rdg.

Visual Inspection OK  
Make Mac Master  
Model TW 50 R  
Class A

Force Set on Torque Wrench in Nm	Reading Observed By Torque Wrench calibration system in Nm	Absolute Error In Nm	Expanded Uncertainty (±)
24	22.94	1.06	2.58 % rdg.
44	42.25	1.75	2.58 % rdg.
68	66.21	1.79	2.58 % rdg.

### Remarks:

- Averages of minimum five readings are reported.
- Suggested due date is given based on customer requirements.
- Calibration points are given based on customer requirements.
- These results are obtained at the time of calibration.
- Any hand written corrections (except @ marked) or photocopies of the report invalidates this certificate.
- The uncertainties are for a confidence probability of not less than 95% with coverage factor k = 2.
- Environment condition during calibration: 23 ± 2°C, 40 to 60% Rh.
- Reference standard no.: IS/ISO 6789.
- Location of performance of calibration → At Lab.
- Condition of instrument found satisfactory during receipt.
- Reference calibration method no.: NCQC/CM/M/052.
- Our masters are directly calibrated through NABL accredited calibration laboratory having direct Traceability with national / international standard.

### Details of master Instrument Used for Calibration

Nomenclature	Make / Model	Id. No. / Sr. No.	Calibration Due Date
Torque Wrench calibration system	Sushma / TS-103F & TDU-RB-103	TS13-0089,0091,0092,0093, / DU13-0032 (NCQC-M/022)	30-03-2023

NCQC System Certificate No. Certificate no. & traceability of master with National Standards

152, 152/2, 152/3, 152/4

→ Our master torque wrench calibration system is calibrated and traceable to National Standard through NABL accredited laboratory Sushma Industries calibration centre, Certificate no. SCPL/CC/2876, 2877, 2878, 2879 /03/2020-2021, Date - 31-03-2021.

(R)

23/12/2021  
Sushma Industries

Traceable To National / International Standards.

Calibrated By

Jaydeep Khatri

Reviewed & Approved By

Jigar Ranchal

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"





Since 1998

# NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058  
Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370  
E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com

Visit our Web Site : [www.calibrationlaboratory.in](http://www.calibrationlaboratory.in)



CC-2128

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

## Calibration Certificate

Name of Customer → **Plasma & Vacuum Technologies**  
Plot No.17, 1/A Road, GIDC, Kathwada,  
Ahmedabad-382 430, Gujarat.

Certificate No. NCQC-M/161221/19  
Date of Issue 17-12-2021  
Date of Calibration 16-12-2021  
Suggested Due Date 15-12-2022

Date Of Receipt / Ref. No. → 16-12-2021

F/CR/M/055, Issue No.01  
Page 1 of 1

ULR – CC212821000005789F

Discipline → **Mechanical Calibration,  
Pressure and Vacuum**

Details of Observation of Unit  
Under Calibration

Identification No. : **PVT/SD/GP70D/16**  
Serial No. : **H 180807**  
Name of Instrument : **Pressure Gauge**

Range 0 – 70 kg/cm<sup>2</sup>  
Resolution 1 kg/cm<sup>2</sup>  
Make / Model LA / EN837-1  
Accuracy =====

Initial Error Nil  
Visual Inspection Ok  
Location =====

Set Pressure on UUC	Reading Observed Master instrument (Xi)		Absolute Error  Xi - Xt		% Error of F.S.  Xi-Xt  / F.S. * 100		Expanded Uncertainty (±) in kg/cm <sup>2</sup>
(Xt)	Increasing Order	Decreasing Order	Increasing Order	Decreasing Order	Increasing Order	Decreasing Order	
0	0.000	0.000	0.000	0.000	0.00%	0.00%	0.292
7	7.112	7.092	0.112	0.092	0.16%	0.13%	0.301
12	12.222	12.233	0.222	0.233	0.32%	0.33%	0.299
21	21.391	21.383	0.391	0.383	0.56%	0.55%	0.298
28	28.335	28.315	0.335	0.315	0.48%	0.45%	0.302
42	42.523	42.534	0.523	0.534	0.75%	0.76%	0.299
56	56.761	56.783	0.761	0.783	1.09%	1.12%	0.303
60	60.638	60.615	0.638	0.615	0.91%	0.88%	0.296
70	70.886	70.865	0.886	0.865	1.27%	1.24%	0.303

### Remarks:

- Averages of minimum three readings are reported.
- Suggested due date is given based on customer requirements.
- These results are obtained at the time of calibration.
- Pressure gauge was calibrated in Hydraulic mode.
- Condition of instrument found satisfactory during receipt.
- Any hand written corrections (except @) or photocopies of the report invalidates this certificate.
- Location of Performance of Calibration → At Lab.
- Environment condition during calibration: 23 ± 1.5°C, 40 to 60% Rh.
- The uncertainties are for a confidence probability of not less than 95% with coverage factor k = 2
- Reference Standard no. : DKD-R 6-1.
- Reference Calibration method no. : NCQC/CM/M/055.
- Our masters are directly calibrated through NABL accredited calibration laboratory having direct traceability with National / international standard.

### Details of Master Instrument Used for Calibration

Nomenclature	Make / Model	Id. No. / Sr. No.	Calibration Due Date
Digital Pressure Gauge	Vijay Enterprises/ MGA N32-2	NCQC/M-146/NVEM1806034	29-06-2022

NCQC System Certificate No.

Certificate no. & Traceability of master with National Standards

299

- Our master Digital Pressure gauge is calibrated and traceable to National Standard through NABL Accredited Laboratory, Vijay Instrumentation Services, Certificate No. VIS/21-22/M-517, Date -29-06-2021.

Traceable To National / International Standards.

Calibrated By

*Mahesh Desai*

Mahesh Desai

Reviewed & Approved By

*Vijay Panchal*



Vijay Panchal

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"



## CALIBRATION CERTIFICATE

Service Request No. :- 2021/07/285		Certificate No. :- HTC/2021/07/11360	
ULR No. :- CC247821000011360F		Certificate Date of Issue :- 28-Jul-2021	
Date of Calibration 22-Jul-2021	Recom. Due Date :- 21-Jul-2022	Discipline Thermal - Temperature	NABL Certificate Due on 05-Dec-2021
Page 1 of 3			
1. Customer :- Plasma & Vacuum Technologies Plot No. 17, Road No. 1A Kathwada GIDC, Ahmedabad-382430			
Received Date :- 21-07-2021	Work Instruction No. :- HTC/WI/10	Environment Condition	
Location of Calibration :- At Lab	Reference Standard :-	Temp. °C	RH %
Condition of Item :- Good	NABL 129 & Euromat og-8	24.4	55.3
2. Description of Item			
Name :- Thermocouple	Range :- 0 to 1200 °C		
ID No. :- PVT/SD/TST/28	L.C. :- ---		
Sr. No. :- ---	Accuracy :- ---		
Make :- TST	Working Range :- Full		
Model / Type :- --- / K - Type	Location :- PVT/SD/DLD/03		
3. Detail of Master equipment used for calibration			
Name	Make/I.D No.	Certificate No.	Certified By
R-type Thermocouple	HTC-EQP-018	NI/2008/029/001	Nishitronic CC-2294
Pt-100 Sensor (4-Wire)	HTC-EQP-090	HTC/2020/08/13373	HTC, & CC-2478
Pt-100 Sensor (4-Wire)	HTC-EQP-091	NI/2008/029/002	Nishitronic CC-2294
6 1/2 Digital Precision Multimeter	HTC-EQP-017	HTC/2021/04/6227	HTC & CC-2478
Cal. Validity			
17-Aug-2022			
13-Aug-2021			
17-Aug-2022			
28-Apr-2022			
<p>The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor <math>k=2</math>, which corresponds to a coverage probability of approximately 95% for normal distribution</p> <p>Note :</p> <p>1) UUC stands for Unit Under Calibration.</p> <p>2) This certificate refers only to the particular item submitted for calibration</p> <p>3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Hi - Tech Calibration, Vapi".</p> <p>4) The calibration results relate only to the item calibrated reported in the certificate are valid at the time of and under the stated conditions of measurement.</p>			
<p>Ankit C. Patel Calibration Engineer</p> <p><i>(Signature)</i> Calibrated By</p>		<p>Dhamesh R. Purohit Quality Manager</p> <p><i>(Signature)</i> Authorised Signatory</p>	





HF-31/4





For Quality Command



Certy No. CC-2478

# HI-TECH CALIBRATION

Head Off.: ROYAL INDUSTRIAL HUB, Gala No. 60, N.H.No. 8,  
Nr. Damian Ganga River, Valvada (VAPI), Tal. Umbergaon Dist. Valsad - 396105.  
Email : hitechvapi@yahoo.com / hitechvapi307@gmail.com  
Web : www.hi-techcalibration.com Cell : 9426832487 / 9662980366

Name	Make/I.D No.	Certificate No.	Certified By	Page 2 of 3
------	--------------	-----------------	--------------	-------------

All Calibration done in SI units and are traceable to National / International standards as per required ISO/IEC/17025

4. Tracibility :
- 1 R-Type Thermocouple Calibrate through NABL Lab Nishitronic CC-2294, Vide Certificate No.NI/2008/029/001, Calibrated on 18-Aug-2020 Traceable to National Standard.
  - 2 Pt-100 Sensor (4-Wire) Calibrate through NABL Lab HTC & CC-2478, Vide Certificate No.: HTC/2020/08/13373, Calibrated on 14-Aug-2020, Traceable to National Standard.
  - 3 Pt-100 Sensor (4-Wire) Calibrate through NABL Lab Nishitronic CC-2294, Vide Certificate No.NI/2008/029/002 Calibrated on 18-Aug-2020, Traceable to National standard.
  - 4 6 1/2 Digital Precision Multimeter Calibrated through NABL Lab HTC & CC-2478, Vied, Certificate No. HTC/2021/04/6227, Calibrated on 20-April-2022, Traceable to National Standard.

(R) 23/02/2022  
Shakil's

The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor  $k=2$ , which corresponds to a coverage probability of approximately 95% for normal distribution.

## Note :

- 1) UUC stands for Unit Under Calibration.
- 2) This certificate refers only to the particular item submitted for calibration
- 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Hi - Tech Calibration,Vapi".
- 4) The calibration results relate only to the item calibrated reported in the certificate are valid at the time of and under the stated conditions of measurement.

Ankit C Patel  
Calibration Engineer

Calibrated By



Dhamesh R. Purohit  
Quality Manager

Authorised Signatory

HF-31/4



Certificate No. :-	HTC/2021/07/11360	Date of Calibration :-	22-Jul-2021	Page
I.D. No. :-	PVT/SD/TST/28	Recom. Due Date :-	21-Jul-2022	3 of 3
ULR No. :-	CC247821000011360F	Discipline :-	Thermal Temperature	

## 5. Calibration Method

Actual temperature reading indicated by UUC is compared with specified stable temperature for a given temperature measured using standard – pt-100 Sensor or Thermocouple & DMM or calibrator.

## 6. Calibration Results :

### A) INSTRUMENTAL ERROR FOR TEMPERATURE

Sr. No.	Cal. Point in °C	UUC Reading in °C	Standard Reading in °C	Error in °C	+/- Expanded Uncertainty in °C
1	19.9	19.9	20.024	-0.124	0.820
2	260.1	260.1	260.275	-0.175	0.650
3	750.2	750.2	750.462	-0.262	2.770
4	1000.4	1000.4	1000.683	-0.283	2.530

Ankit C Patel


Calibration Engineer

*Acad*  
Calibrated By

Kavita M. Panwala

Verification Engineer

*KMP*  
Checked By



Dhamesh R. Purohit

Quality Manager

*Shakti S*  
Authorised Signatory

HF-31C/00



## CALIBRATION CERTIFICATE

Service Request No. :- 2021/07/285		Certificate No. :- HTC/2021/07/11361	
ULR No. :- CC247821000011361F		Certificate Date of Issue :- 28-Jul-2021	
Date of Calibration 22-Jul-2021	Recom.Due Date :- 21-Jul-2022	Discipline Thermal - Temperature	NABL Certificate Due on 05-Dec-2021
Page 1 of 3			
1. Customer :- Plasma & Vacuum Technologies Plot No. 17, Road No. 1A Kathwada GIDC, Ahmedabad-382430			
Received Date :- 21-07-2021	Work Instruction No. :- HTC/WI/10	Environment Condition	
Location of Calibration :- At Lab	Reference Standard :-	Temp. °C	RH %
Condition of Item :- Good	NABL 129 & Euromat cg-8	25.3	55.0
2. Description of Item			
Name :- Thermocouple	Range :- 0 to 1200 °C		
ID No. :- PVT/SD/TST/29	L.C. :- ---		
Sr. No. :- ---	Accuracy :- ---		
Make :- TST	Working Range :- Full		
Model / Type :- --- / K - Type	Location :- PVT/SD/DLD/03		
3. Detail of Master equipment used for calibration			
Name	Make/I.D No.	Certificate No.	Certified By
R-type Thermocouple	HTC-EQP-018	NI/2008/029/001	Nishitronic CC-2294
Pt-100 Sensor (4-Wire)	HTC-EQP-090	HTC/2020/08/13373	HTC, & CC-2478
Pt-100 Sensor (4-Wire)	HTC-EQP-091	NI/2008/029/002	Nishitronic CC-2294
6 1/2 Digital Precision Multimeter	HTC-EQP-017	HTC/2021/04/6227	HTC & CC-2478
Cal. Validity			
17-Aug-2022			
13-Aug-2021			
17-Aug-2022			
28-Apr-2022			
<p>The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor <math>k=2</math>, which corresponds to a coverage probability of approximately 95% for normal distribution.</p> <p>Note :</p> <p>1) UUC stands for Unit Under Calibration.</p> <p>2) This certificate refers only to the particular item submitted for calibration</p> <p>3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Hi - Tech Calibration,Vapi"</p> <p>4) The calibration results relate only to the item calibrated reported in the certificate are valid at the time of and under the stated conditions of measurement</p>			
<p>Ankit C Patel Calibration Engineer</p> <p><i>Ankit C Patel</i> Calibrated By</p>		<p><i>Shakil S.</i></p> <p>Dharmesh R. Purohit Quality Manager</p> <p><i>Dharmesh R. Purohit</i> Authorised Signatory</p>	



HF-31/4



Name	Make/I.D No.	Certificate No.	Certified By	Page 2 of 3
------	--------------	-----------------	--------------	-------------

All Calibration done in SI units and are traceable to National / International standards as per required ISO/IEC/17025

#### 4. Tracibility :

- 1 R-Type Thermocouple Calibrate through NABL Lab Nishitronic CC-2294, Vide Certificate No.NI/2008/023/001. Calibrated on 18-Aug-2020 Traceable to National Standard.
- 2 Pt-100 Sensor (4-Wire) Calibrate through NABL Lab HTC & CC-2478, Vide Certificate No.: HTC/2020/08/13373. Calibrated on 14-Aug-2020, Traceable to National Standard.
- 3 Pt-100 Sensor (4-Wire) Calibrate through NABL Lab Nishitronic CC-2294, Vide Certificate No.NI/2008/020/002 Calibrated on 18-Aug-2020, Traceable to National standard.
- 4 6 1/2 Digital Precision Multimeter Calibrated through NABL Lab HTC & CC-2478, Vide Certificate No.: HTC/2021/04/6227. Calibrated on 20-April-2022, Traceable to National Standard.

The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor  $k=2$ , which corresponds to a coverage probability of approximately 95% for normal distribution.

#### Note :

- 1) UUC stands for Unit Under Calibration.
- 2) This certificate refers only to the particular item submitted for calibration
- 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Hi - Tech Calibration, Vapi"
- 4) The calibration results relate only to the item calibrated reported in the certificate are valid at the time of and under the stated conditions of measurement.


Ankit C Patel  
Calibration Engineer

*Signature of Ankit C Patel*  
Calibrated By



Dharmesh R. Purohit  
Quality Manager

*Signature of Dharmesh R. Purohit*  
Authorised Signatory



HF-31/4



Certificate No. :-	HTC/2021/07/11361	Date of Calibration :-	22-Jul-2021	Page
I.D. No. :-	PVT/SD/TST/29	Recom. Due Date :-	21-Jul-2022	3 of 3
ULR No. :-	CC247821000011361F	Discipline :-	Thermal - Temperature	

## 5. Calibration Method

Actual temperature reading indicated by UUC is compared with specified stable temperature for a given temperature measured using standard – pt-100 Sensor or Thermocouple & DMM or calibrator.

## 6. Calibration Results :

### A) INSTRUMENTAL ERROR FOR TEMPERATURE

Sr. No.	Cal. Point in °C	UUC Reading in °C	Standard Reading in °C	Error in °C	+/- Expanded Uncertainty in °C
1	20.0	20.0	20.142	-0.142	0.820
2	260.1	260.1	260.275	-0.175	0.650
3	750.3	750.3	750.587	-0.287	2.770
4	1000.5	1000.5	1000.852	-0.352	2.530

Ankit C Patel

Calibration Engineer

*AC Patel*  
Calibrated By

Kavita M. Panwala

Verification Engineer

*KMP*  
Checked By



Dharmesh R. Purohit

Quality Manager

*Dharmesh R. Purohit*  
Authorised Signatory



23/02/2022  
*Shahat S.*

HF-31C/00