

**MANUFACTURING :**

Custom - built Plasma &amp; Vacuum Components and Assemblies

**SERVICES :**

Helium Leak Testing, Methane Leak Testing as per API &amp; ISO-15848 Standards

Vacuum Brazing, Vacuum Heat Treatment, Calibration of Vacuum Gauges



Representative of Saes Getters, Italy for Vacuum Pumps in India

Plot No. 17, Road 1 - A,

GIDC, Kathwada, Ahmedabad-382430.

Phone : +91 79-29705319, Fax : 079-29709216

Email : [pvt@plasvac.com](mailto:pvt@plasvac.com)Website : [www.plasvac.com](http://www.plasvac.com)

Certificate No : 624202202276 ✓

Date : February 27, 2022 ✓

**API STANDARD 624 TEST CERTIFICATE****Type Testing of Rising Stem Valves Equipped with Graphite Packing For Fugitive Emissions**

This is to certify that the below mentioned valve has been inspected by Plasma & Vacuum Technologies and found to be satisfactory in accordance with the requirements of API 624, First Edition 2014.

Manufacturer:	Fluid-O-Mech Controls Inc. ✓
Address:	14/1A, Pancharatna Industrial Estate, Near Ramol Bridge, Vatva G.I.D.C., Ahmedabad – 382 445. Gujarat. India.
Location of Test:	Plasma & Vacuum Technologies, Plot No.17, Road 1-A, GIDC Kathwada, Ahmedabad 382430, India.

Product :	Valve Details:	Gate Valve, 12" 300# ✓
	Design Standard:	API 600 ✓
Packing :	Make: New Empire Gaskets, Model: Seal-It.	
Body/Bonnet Material :	ASTM A216 Gr. WCB ✓	
Stem Material :	ASTM A276 TP410 ✓	
GA Drawing No:	FMC/300X300#/GTV/SW/00 ✓	
Test Fluid :	Methane (>97 % Purity) ✓	
Test Pressure :	41.4 barG (~600 psiG) ✓	
Test Temperature :	At Ambient and 260°C (~ 500°F) ✓	
Mechanical Cycle :	310 Cycles over 3 Thermal Cycles ✓	
Maximum Leakage :	46.8	ppmv
Average Leakage :	25.7	ppmv
Acceptance Criteria :	< 100	ppmv maximum
Test Result :	PASS	
Test Report No :	PVT/SD/FET/R/202202/27 ✓	
Test Start Date :	February 23, 2022 ✓	
Test End Date :	February 25, 2022 ✓	

Valve Qualified : All valves of the same basic design as the test valve may be deemed to have been type tested, subject to the following additional limitation. For API 600 Valves, 12" (DN 300) Class 300 test valve qualifies all valves from 8" (DN 200) through 14" (DN 350) in pressure class 150 and 300.✓


Test Conducted by

Venkat N. Ramani  
ASNT Level III (LT) 183918  
Plasma & Vacuum Technologies



Test Witnessed By

*Witnessed*  
*27/02/2022*  
*Shantir*





## MANUFACTURING :

Custom - built Plasma & Vacuum Components and Assemblies

## SERVICES :

Helium Leak Testing, Methane Leak Testing as per API & ISO-15848 Standards

Vacuum Brazing, Vacuum Heat Treatment, Calibration of Vacuum Gauges

Saes  
getters

Representative of Saes Getters, Italy for Vacuum Pumps in India

Plot No. 17, Road 1 - A,

GIDC, Kathwada, Ahmedabad-382430.

Phone : +91 79-29705319, Fax : 079-29709216

Email : [pvt@plasvac.com](mailto:pvt@plasvac.com)

Website : [www.plasvac.com](http://www.plasvac.com)

## API STANDARD 624 FUGITIVE EMISSIONS TEST REPORT

F/CR/10.2  
RevNo.00

Report No. :	PVT/SD/FET/R/202202/27 ✓	Report Date :	27 February 2022 ✓
Manufacturer :	Fluid-O-Mech Controls Inc. ✓	Valve Stem Type :	Rising Stem ✓
Valve - Type, Size, Class :	Gate Valve, 12" 300# ✓	API / ASME Design Standard:	API 600 ✓
Packing Description :	Number of Rings : 6 nos OD : 60.8 mm, ID : 41.5 mm Thickness : 9.5 mm Stack Height : 57 mm Stuffing box depth : 60 mm		
Make: New Empire Gaskets, Model: Seal-It.	Body / Bonnet Gasket Connection: SPW SS316 + Graphite Filler.		
Packing Manufacturer :	New Empire Gaskets		
	Gland Follower Exposed Length: 29mm	Body-Bonnet connection : Bolted.	
Gland Nut recommended torque: 140Nm	Gland Follower Insertion Depth: 3mm	Body/Bonnet Nut recommended torque: 510Nm	
Test Start Date :	23 February 2022 ✓	Valve Selection : by Manufacturer; Selection Date: 17.02.2022 ✓	
Test Completion Date :	25 February 2022 ✓	Selected by : QA/QC Dept. Fluid-O-Mech Controls Inc. Ahmedabad.	

## SUMMARY OF FUGITIVE EMISSIONS TEST DATA

Test Segment	Cycle	Methane Pressure (barG)	Temperature at Body (°C)	Temperature at Stem (°C)	Static Leak Measurement (ppmv)	Dynamic Leak Measurement (ppmv)	Remarks
Ambient Temperature T(a) = RT, P(a)=41.4 barG	0	41.4	26	26 ✓	12.6	16.7	OK
	50	41.4	27	27	13.7		
Elevated Temperature T(e) = 260°C, P(e) = 41.4 barG	51	41.4	264	265	15.0	26.1	OK
	100	41.4	266	268	21.6		
Ambient Temperature T(a) = RT, P(a)=41.4 barG	101	41.4	30	31	18.3	30.9	OK
	150	41.4	33	33	24.8		
Elevated Temperature T(e) = 260°C, P(e) = 41.4 barG	151	41.4	262	260	18.8	34.7	OK
	200	41.4	262	260	30.8		
Ambient Temperature T(a) = RT, P(a)=41.4 barG	201	41.4	25	25	16.8	37.0	OK
	250	41.4	27	27	29.0		
Elevated Temperature T(e) = 260°C, P(e) = 41.4 barG	251	41.4	261	266 ✓	25.9	46.8	OK
	300	41.4	263	267	36.9		
Ambient Temperature T(a) = RT, P(a)=41.4 barG	301	41.4	40	40	18.0	34.4	OK
	310	41.4	40	40	31.5		

Valve Serial Number: FMCI-118 (Heat No.: Body: G55, Bonnet: G55) ✓

Running Torque - First Cycle : 32 Nm, Last cycle : 28 Nm

FINAL REMARK : Pass ( ✓ )

Pre-test Preparations & Adjustments : 1) Gland bolting, Body/Bonnet and Valve open-close torques were verified; 2) The air in the valve cavity was evacuated and purged with Helium prior to starting the testing; 3) The stem orientation was kept vertical; 4) External valve heating was designed, with heating coils and wrapped by Alumina wool; 5) Electrical Operation using wheel was set for cycling (opening and closing) the valves.


Notes : 1) The test valve was randomly selected; 2) The test medium was methane with purity > 97%; 3) The test was conducted in a safe, well ventilated and protected environment; 4) No packing adjustment was done during test. Running torques at the start and end were measured and recorded; 5) Methane Detector (Model TVA 2020, Thermofisher) was used for ambient monitoring; 6) Methane Detector (Model TVA 2020, Thermofisher) was used for methane leakage at gland and body / bonnet joint. The Probe was calibrated prior to each measurement using an external calibration gas with known Methane concentration. T(a)= Ambient Temperature, T(e)= Elevated Temperature, P(a)= Ambient Pressure & P(e)= Elevated Pressure, RT=Room Temperature

Test Witnessed by

Test Conducted by

Test Witnessed By

Venkat N. Ramani  
ASNT Level III (LT)








## Annexure 1 : SET UP

Report No.: PVT/SD/FET/R/202202/ 27


Report Date: 27 February 2022



- 1 : Methane Gas Cylinder; 2 : Isolation Valve; 3 : Pressure Gauge;  
4 : Valve under Test; 5 : Temperature Sensor; 6 : Hot Zone;  
7 : Methane Leak Detector; 8 : Sniffer Probe; 9 : Valve operation-Rising Stem.



Plot No. 17, Road 1 - A,  
GIDC, Kathwada, Ahmedabad-382430.  
Phone : +91 79-29705319, Fax : 079-29709216  
Email : [pvt@plasvac.com](mailto:pvt@plasvac.com)  
Website : [www.plasvac.com](http://www.plasvac.com)





## ANNEXURE 3 : TEST REPORT - HELIUM LEAK TESTING

Report No :	PVT/SD/FET- HLT/R/202202-27 ✓		Report Date :	February 23, 2022 ✓		
Client :	Fluid-O-Mech Controls Inc.		Testing Equipment :	HLD ASM 310 ADIXEN		
Manufacturer :	Fluid-O-Mech Controls Inc.		Calibrated Leak(s) :	PVT/CD/SL/03, PVT/CD/SL/04		
Item :	Gate Valve, 12" 300# ✓		Ref. Code(s) :	ASME Sec V, Art 10, Detector Probe Technique		
Reference:	Prior to API624 Test of valve		Qualification:	Helium Leakage Rate : <1.0x10 <sup>-6</sup> mbar l/s at Body-Bonnet and Gland joints at 41.4 barG pressure (10% of Helium + 90% Nitrogen).✓		
<b>Leak Detector and Probe Calibration</b>						
	Standard Leak Value		Observed Leak Value		Instrument Sensitivity	
A	2.48 x 10 <sup>-8</sup>	mbar l/s	~ 2.5 x 10 <sup>-8</sup>	mbar l/s	better than 10 <sup>-9</sup>	std cc/s
B	1.13 x 10 <sup>-6</sup>	mbar l/s	~ 1.1 x 10 <sup>-6</sup>	mbar l/s	CF = 1.0	Response < 5 s
	Stage	Test Date	Body		Gland Joint	Remarks
			Leak Rate (mbar l/s)		Leak Rate (mbar l/s)	
1	Prior to API624 Test	Feb 23, 2022 ✓	5.7 x 10 <sup>-7</sup> ✓		7.2 x 10 <sup>-7</sup> ✓	Leak Tightness is OK
Test Witnessed by			Test Conducted by		Test Witnessed By	
Fluid-O-Mech Controls Inc.			Venkat N. Ramani ASNT Level III (LT) Plasma&Vacuum Technologies		 	



## Annexure 4 CALIBRATION DATA AND METHANE LEAKAGE EVALUATIONS

Report No.: **PVT/SD/FET/R/202202/27**

Report Date: **27 February 2022**

Test Segment	Cycle	Condition	Ambient Leak Rate Value (ppmv)	Calibration Leak Rate Value (ppmv)	Detector Sensitivity	Observed Leak Rate (ppmv)	Measured Leak Rate (ppmv)
Ambient Temperature T(a) = RT, P(a)=41.4 barG	0	Static	2.3	75.4	1.02	14.7	12.6
	50	Static	4.2		1.05	17.3	13.7
	50	Dynamic				20.2	16.7
Elevated Temperature T(e) =260°C, P(e) = 41.4barG	51	Static	5.1	75.8	1.05	19.3	15.0
	100	Static	7.6		1.09	27.4	21.6
	100	Dynamic				31.5	26.1
Ambient Temperature T(a) = RT, P(a)=41.4 barG	101	Static	2.8	76.3	1.01	20.8	18.3
	150	Static	5.2		1.05	28.9	24.8
	150	Dynamic				34.7	30.9
Elevated Temperature T(e) =260°C, P(e) = 41.4barG	151	Static	8.0	76.7	1.08	25.3	18.8
	200	Static	3.4		1.02	33.7	30.8
	200	Dynamic				37.5	34.7
Ambient Temperature T(a) = RT, P(a)=41.4 barG	201	Static	7.0	74.9	1.10	22.3	16.8
	250	Static	2.1		1.02	30.4	29.0
	250	Dynamic				38.2	37.0
Elevated Temperature T(e) =260°C, P(e) = 41.4barG	251	Static	7.7	75.4	1.10	31.2	25.9
	300	Static	5.9		1.07	40.3	36.9
	300	Dynamic				49.5	46.8
Ambient Temperature T(a) = RT, P(a)=41.4 barG	301	Static	8.4	75.9	1.10	24.7	18.0
	310	Static	2.8		1.02	33.7	31.5
	310	Dynamic				36.5	34.4
Average Value:							25.7

Calibrated Leak used (ml/m) : SL	0.072
Flow Rate at Methane Leak Detector (lpm) : FR	0.966
Estimated PPM value (PPMv) = (SL/FR)x1000	74.53

### Leakage at Body-Bonnet Connection

Cycle Number	Body Temperature (°C)	Pressure (barG)	Ambient Leak Rate Value (ppmv)	Calibration Leak Rate Value (ppmv)	Detector Sensitivity	Observed Leak Rate Value (ppmv)	Measured Leak Rate Value (ppmv)
0	28	41.4	2.3	75.4	1.02	12.3	10.2
310	40	41.4	2.8	75.9	1.02	24.5	22.1

Measured Leak Rate = (Observed Leak Rate – Ambient Leak Rate Value) x  
[Estimated PPM value / (Calibration Leak Rate Value – Ambient Leak Rate Value)]

Test Witnessed by  Fluid-O-Mech Controls Inc.	Test Conducted by  Venkat N. Ramani ASNT Level III (LT) Plasma&Vacuum Technologies	Test Witnessed By   27/02/2022 Shariat
---	--	--





## MANUFACTURING :

Custom - built Plasma & Vacuum Components and Assemblies

## SERVICES :

Helium Leak Testing, Methane Leak Testing as per API & ISO-15848 Standards

Vacuum Brazing, Vacuum Heat Treatment, Calibration of Vacuum Gauges

**Saes**  
getters  
Representative of Saes Getters, Italy for Vacuum Pumps in India

Plot No. 17, Road 1 - A,  
GIDC, Kathwada, Ahmedabad-382430.  
Phone : +91 79-29705319, Fax : 079-29709216  
Email : [pvt@plasvac.com](mailto:pvt@plasvac.com)  
Website : [www.plasvac.com](http://www.plasvac.com)

## Annexure 5


### Methane Leakage Measurements

Report No. : PVT/SD/FET/R/202202/ 27


Report Date: 27 February 2022




(a)



(b)



(c)



(d)

Methane leakage at : a) Ambient, Static, 0 cycles; b) Ambient, Static, 101 cycles; c) Ambient, Static, 201 Cycles ;  
d) Elevated, Static, 300 Cycles ;



W





# Plasma And Vacuum Technologies



## MANUFACTURING :

Custom - built Plasma & Vacuum Components and Assemblies

## SERVICES :

Helium Leak Testing, Methane Leak Testing as per API & ISO-15848 Standards

Vacuum Brazing, Vacuum Heat Treatment, Calibration of Vacuum Gauges

Saes  
getters

Representative of Saes Getters, Italy for Vacuum Pumps in India

Plot No. 17, Road 1 - A,  
GIDC, Kathwada, Ahmedabad-382430.

Phone : +91 79-29705319, Fax : 079-29709216



Email : [pvt@plasvac.com](mailto:pvt@plasvac.com)

Website : [www.plasvac.com](http://www.plasvac.com)


## Annexure 6 Temperature Measurements

Report No. : PVT/SD/FET/R/202202/ 27

Report Date: 27 February 2022



12 Inch 300# Gate Valve ; Fluid-O-Mech Controls Inc.  
API 624 Test : Temperature Profile ✓



2





## MANUFACTURING :

Custom - built Plasma & Vacuum Components and Assemblies

## SERVICES :

Helium Leak Testing, Methane Leak Testing as per API & ISO-15848 Standards

Vacuum Brazing, Vacuum Heat Treatment, Calibration of Vacuum Gauges

**Saes** getters Representative of Saes Getters, Italy for Vacuum Pumps in India

Plot No. 17, Road 1 - A,  
GIDC, Kathwada, Ahmedabad-382430.

Phone : +91 79-29705319, Fax : 079-29709216


Email : [pvt@plasvac.com](mailto:pvt@plasvac.com)

Website : [www.plasvac.com](http://www.plasvac.com)


## Annexure 7 Pressure Measurements

Report Date: PVT/SD/FET/R/202202/ 27

Report Date: 27 February 2022



12 Inch 300# Gate Valve ; Fluid-O-Mech Controls Inc.  
API 624 Test : Pressure Profile





## MANUFACTURING :

Custom - built Plasma & Vacuum Components and Assemblies

## SERVICES :

Helium Leak Testing, Methane Leak Testing as per API & ISO-15848 Standards

Vacuum Brazing, Vacuum Heat Treatment, Calibration of Vacuum Gauges

**Saes** getters Representative of Saes Getters, Italy for Vacuum Pumps in India

Plot No. 17, Road 1 - A,  
GIDC, Kathwada, Ahmedabad-382430.  
Phone : +91 79-29705319, Fax : 079-29709216  
Email : [pvt@plasvac.com](mailto:pvt@plasvac.com)  
Website : [www.plasvac.com](http://www.plasvac.com)


## Annexure 8 Valve Operation Mechanism

Report No.: PVT/SD/FET/R/202202/ 27

Report Date: 27 February 2022



API 624 Test : Methane Leakage Measurement Points





## MANUFACTURING :

Custom - built Plasma & Vacuum Components and Assemblies

## SERVICES :

Helium Leak Testing, Methane Leak Testing as per API & ISO-15848 Standards

Vacuum Brazing, Vacuum Heat Treatment, Calibration of Vacuum Gauges

**Saes**  
getters  
Representative of Saes Getters, Italy for Vacuum Pumps in India




Plot No. 17, Road 1 - A,  
GIDC, Kathwada, Ahmedabad-382430.  
Phone : +91 79-29705319, Fax : 079-29709216  
Email : [pvt@plasvac.com](mailto:pvt@plasvac.com)  
Website : [www.plasvac.com](http://www.plasvac.com)

## Annexure 9

### Post – Test Valve Opened Up Photographs

Report No.: PVT/SD/FET/R/202202/ 27

Report Date: 27 February 2022



**Remarks:** The disassembled Valve parts were inspected and the condition was found satisfactory.



## List of Calibration Reports of Instruments / Documents attached

Report No.: PVT/SD/FET/R/202202/ 27

Report Date: 27 February 2022

- 1 Methane Gas Cylinder 709
- 2 Gas Monitor Pump (Flow Rate), Report No. PVT/SD/FRM/R/2022/01
- 3 Standard Leak, SI No PVT/TD/CL/06, Report No. PVT/SD/FRM/R/2022/02
- 4 Helium Standard Leak, SI No. PVT/CD/SL/03
- 5 Helium Standard Leak, SI No. PVT/CD/SL/04
- 6 Torque Wrench, MACMASTER, PVT/SD/TW/01 & 02
- 7 Pressure Gauge, LA, PVT/SD/GP70D/22 ✓
- 8 Temperature Sensor, Thermal Sense Tech, PVT/SD/TST/31 & 32 ✓
- 9 ASNT NDT Level III certification of V N Ramani
- 10 API 622 Qualified Packing Certificate

Page 12 of 12



verified  
27/02/2022  
Shakil S.







Since 1998

# NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058  
Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370

E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com

Visit our Web Site : [www.calibrationlaboratory.in](http://www.calibrationlaboratory.in)



CC-2128

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

## Calibration Certificate

Name of Customer → **Plasma & Vacuum Technologies**  
Plot No.17, 1/A Road, GIDC Kathwada,  
Ahmedabad-382 430, Gujarat.

Certificate No. NCQC-M/161221/03  
Date of Issue 17-12-2021  
Date of Calibration 16-12-2021  
Suggested Due Date 15-12-2022

Date Of Receipt / Ref. No. → 16-12-2021

F/CR/M/052, Issue No.01

Page 1 of 1

ULR – CC212821000005773F

Discipline → Mechanical Calibration,  
Force, Torque

Details of Observation of Unit Under Calibration

Identification No. : PVT/SD/TW/01  
Serial No. : 24B – 41  
Name of Instrument : Torque Wrench

Range	70 – 340 Nm	Visual Inspection	OK
Least Count	10 Nm	Make	Mac Master
Type	Type II	Model	TW 250 R
Accuracy	± 4 % of rdg.	Class	A

Force Set on Torque Wrench in Nm	Reading Observed By Torque Wrench calibration system in Nm	Absolute Error In Nm	Expanded Uncertainty (±)
110	107.9	2.1	2.42 % rdg.
230	226.7	3.3	2.42 % rdg.
320	324.3	4.3	2.42 % rdg.

### Remarks:

- Averages of minimum five readings are reported.
- Suggested due date is given based on customer requirements.
- Calibration points are given based on customer requirements.
- These results are obtained at the time of calibration.
- Any hand written corrections (except @ marked) or photocopies of the report invalidates this certificate.
- The uncertainties are for a confidence probability of not less than 95% with coverage factor k = 2.
- Environment condition during calibration: 23 ± 2 °C, 40 to 60% Rh.
- Reference standard no.: IS/ISO 6789.
- Location of performance of calibration → At Lab.
- Condition of instrument found satisfactory during receipt.
- Reference calibration method no.: NCQC/CM/M/052.
- Our masters are directly calibrated through NABL accredited calibration laboratory having direct Traceability with national / international standard.

### Details of master Instrument Used for Calibration

Nomenclature	Make / Model	Id. No. / Sr. No.	Calibration Due Date
Torque Wrench calibration system	Sushma / TS-103F & TDU-RB-103	TS13-0089,0091,0092,0093, / DU13-0032 (NCQC-M/022)	30-03-2023

NCQC System Certificate No. 152, 152/2, 152/3, 152/4

Certificate no. & traceability of master with National Standards

Our master torque wrench calibration system is calibrated and traceable to National Standard through NABL accredited laboratory Sushma Industries calibration centre, Certificate no. SCPL/CC/2876.2877-2878-2879 /03/2020-2021, Date - 31-03-2021.

(R)

23/12/2021  
Shahid S

Traceable To National / International Standards.

Calibrated By

Jk Khatri  
Jaydeep Khatri

Reviewed & Approved By

Jigar Panchal

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"





# NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058  
Ph. +91-79-29795322, 29795323 • Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370  
E-mail : ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com  
Visit our Web Site : [www.calibrationlaboratory.in](http://www.calibrationlaboratory.in)



CC-2128

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

## Calibration Certificate

Name of Customer → **Plasma & Vacuum Technologies**  
Plot No.17, 1/A Road, GIDC Kathwada,  
Ahmedabad-382 430, Gujarat.

Certificate No. NCQC-M/161221/01  
Date of Issue 17-12-2021  
Date of Calibration 16-12-2021  
Suggested Due Date 15-12-2022

Date Of Receipt / Ref. No. → 16-12-2021

F/CR/M/052, Issue No.01  
Page 1 of 1

ULR – CC212821000005771F

Discipline → Mechanical Calibration,  
Force, Torque

Details of Observation of Unit Under Calibration

Identification No. : PVT/SD/TW/02  
Serial No. : 24R – 12  
Name of Instrument : Torque Wrench

Range 10 – 68 Nm  
Least Count 2 Nm  
Type Type II  
Accuracy  $\pm 4\%$  of rdg.

Visual Inspection OK  
Make Mac Master  
Model TW 50 R  
Class A

Force Set on Torque Wrench in Nm	Reading Observed By Torque Wrench calibration system in Nm	Absolute Error In Nm	Expanded Uncertainty ( $\pm$ )
24	22.94	1.06	2.58 % rdg.
44	42.25	1.75	2.58 % rdg.
68	66.21	1.79	2.58 % rdg.

### Remarks:

- Averages of minimum five readings are reported.
- Suggested due date is given based on customer requirements.
- Calibration points are given based on customer requirements.
- These results are obtained at the time of calibration.
- Any hand written corrections (except @ marked) or photocopies of the report invalidates this certificate.
- The uncertainties are for a confidence probability of not less than 95% with coverage factor  $k = 2$ .
- Environment condition during calibration:  $23 \pm 2^\circ\text{C}$ , 40 to 60% Rh.
- Reference standard no.: IS/ISO 6789.
- Location of performance of calibration → At Lab.
- Condition of instrument found satisfactory during receipt.
- Reference calibration method no.: NCQC/CM/M/052.
- Our masters are directly calibrated through NABL accredited calibration laboratory having direct Traceability with national / international standard.

### Details of master Instrument Used for Calibration

Nomenclature	Make / Model	Id. No. / Sr. No.	Calibration Due Date
Torque Wrench calibration system	Sushma / TS-103F & TDU-RB-103	TS13-0089,0091,0092,0093, / DU13-0032 (NCQC-M/022)	30-03-2023

NCQC System Certificate No.

Certificate no. & traceability of master with National Standards

152, 152/2, 152/3, 152/4

Our master torque wrench calibration system is calibrated and traceable to National Standard through NABL accredited laboratory Sushma Industries calibration centre, Certificate no. SCPL/CC/2876,2877, 2878, 2879 / 03/2020-2021, Date - 31-03-2021.

Traceable To National / International Standards.

Calibrated By

Jaydeep Khatri

Reviewed & Approved By

Jigar Ranchal

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"





# NATIONAL CENTRE FOR QUALITY CALIBRATION

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon-Ambli Road, Ambli, Ahmedabad-380 058

Ph. +91-79-29795322, 29795323 • Fax: +91-79-29795323, Cell No. +91-9327017517, +91-9328616370

E-mail: ncqc@calibrationlaboratory.in, • calibrationlab.ncqc@gmail.com

Visit our Web Site : [www.calibrationlaboratory.in](http://www.calibrationlaboratory.in)



CC-2128

Precision Calibration with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

## Calibration Certificate

Name of Customer → <b>Plasma &amp; Vacuum Technologies</b> Plot No.17, 1/A Road, GIDC, Kathwada, Ahmedabad-382 430, Gujarat.		Certificate No. NCQC-M/161221/09
Date Of Receipt / Ref. No. → 16-12-2021		Date of Issue 17-12-2021
ULR – CC212821000005779F		Date of Calibration 16-12-2021
Discipline → Mechanical Calibration, Pressure and Vacuum		Suggested Due Date 15-12-2022
Details of Observation of Unit Under Calibration		F/CR/M/055, Issue No.01 Page 1 of 1
Identification No. : PVT/SD/GP70D/22		
Serial No. : AQ008		
Name of Instrument : Pressure Gauge		
Range	0 – 70 kg/cm <sup>2</sup>	Initial Error
Resolution	1 kg/cm <sup>2</sup>	Visual Inspection
Make / Model	LA / EN837-1	Location
Accuracy	=====	Nil Ok =====

Set Pressure on UUC	Reading Observed Master instrument (Xi)		Absolute Error  Xi - Xt		% Error of F.S.  Xi-Xt  / F.S. * 100		Expanded Uncertainty (±) in kg/cm <sup>2</sup>
(Xt)	Increasing Order	Decreasing Order	Increasing Order	Decreasing Order	Increasing Order	Decreasing Order	
0	0.000	0.000	0.000	0.000	0.00%	0.00%	0.292
7	7.013	7.011	0.013	0.011	0.02%	0.02%	0.292
12	12.031	12.044	0.031	0.044	0.04%	0.06%	0.293
21	21.050	21.056	0.050	0.056	0.07%	0.08%	0.292
28	28.069	28.066	0.069	0.066	0.10%	0.09%	0.292
42	42.053	42.066	0.053	0.066	0.08%	0.09%	0.293
56	56.084	56.068	0.084	0.068	0.12%	0.10%	0.294
60	60.109	60.097	0.109	0.097	0.16%	0.14%	0.293
70	70.121	70.143	0.121	0.143	0.17%	0.20%	0.296

### Remarks:

- Averages of minimum three readings are reported.
- Suggested due date is given based on customer requirements.
- These results are obtained at the time of calibration.
- Pressure gauge was calibrated in Hydraulic mode.
- Condition of instrument found satisfactory during receipt.
- Any hand written corrections (except @) or photocopies of the report invalidates this certificate.
- Location of Performance of Calibration → At Lab.
- Environment condition during calibration: 23 ± 1.5°C, 40 to 60% Rh.
- The uncertainties are for a confidence probability of not less than 95% with coverage factor k = 2
- Reference Standard no. : DKD-R 6-1.
- Reference Calibration method no. : NCQC/CM/M/055.
- Our masters are directly calibrated through NABL accredited calibration laboratory having direct traceability with National / international standard.

### Details of Master Instrument Used for Calibration

Nomenclature	Make / Model	Id. No. / Sr. No.	Calibration Due Date
Digital Pressure Gauge	Vijay Enterprises/ MGA N32-2	NCQC/M-146/NVEM1806034	29-06-2022
NCQC System Certificate No.	Certificate no. & Traceability of master with National Standards		
299	Our master Digital Pressure gauge is calibrated and traceable to National Standard through NABL Accredited Laboratory, Vijay Instrumentation Services, Certificate No. VIS/21-22/M-517, Date -29-06-2021.		

### Traceable To National / International Standards.

Calibrated By		Mahesh Desai	Reviewed & Approved By		Jigar Panchal
---------------	--	--------------	------------------------	--	---------------

NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"



## CALIBRATION CERTIFICATE

Service Request No. :- 2021/07/285 Certificate No. :- HTC/2021/07/11363

ULR No. :- CC247821000011363F Certificate Date of Issue :- 28-Jul-2021

Date of Calibration 22-Jul-2021	Recom.Due Date 21-Jul-2022	Discipline Thermal - Temperature	NABL Certificate Due on 05-Dec-2021	Page 1 of 3
------------------------------------	-------------------------------	-------------------------------------	--	----------------

1. Customer :- Plasma & Vacuum Technologies  
Plot No. 17, Road No. 1A Kathwada GIDC, Ahmedabad-382430

Received Date 21-07-2021	Work Instruction No. HTC/WI/10	Environment Condition Temp. °C RH % 23.2 55.0
Location of Calibration At Lab	Reference Standard : NABL 129 & Euromat cg-8	
Condition of Item Good		

### 2. Description of Item

Name :- Thermocouple	Range :- 0 to 1200 °C
ID No. :- PVT/SD/TST/31	L.C. :-
Sr. No. :-	Accuracy :-
Make :- TST	Working Range :- Full
Model / Type :- / K - Type	Location :- PVT/SD/DLD/03

### 3. Detail of Master equipment used for calibration

Name	Make/I.D No.	Certificate No.	Certified By	Cal. Validity
R-type Thermocouple	HTC-EQP-018	NI/2008/029/001	Nishitronic CC-2294	17-Aug-2022
Pt-100 Sensor (4-Wire)	HTC-EQP-090	HTC/2020/08/13373	HTC, & CC-2478	13-Aug-2021
PI-100 Sensor (4-Wire)	HTC-EQP-091	NI/2008/029/002	Nishitronic CC-2294	17-Aug-2022
6 1/2 Digital Precision Multimeter	HTC-EQP-017	HTC/2021/04/6227	HTC & CC-2478	28-Apr-2022

The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor  $k=2$ , which corresponds to a coverage probability of approximately 95% for normal distribution.

#### Note :

- 1) UUC stands for Unit Under Calibration.
- 2) This certificate refers only to the particular item submitted for calibration.
- 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Hi - Tech Calibration,Vapi".
- 4) The calibration results relate only to the item calibrated reported in the certificate are valid at the time of and under the stated conditions of measurement.

Ankit C Patel  
Calibration Engineer

Calibrated By



Dharmesh R. Purohit  
Quality Manager

Authorised Signatory

HF-31/4



Name	Make/I.D No.	Certificate No.	Certified By	Page 2 of 3
------	--------------	-----------------	--------------	-------------

All Calibration done in SI units and are traceable to National / international standards as per required ISO/IEC/17025

4. Tracibility :
- 1 R-Type Thermocouple Calibrate through NABL Lab Nishitronic CC-2294, Vide Certificate No NI/2008/029/001. Calibrated on 15-Aug-2020 Traceable to National Standard.
  - 2 Pt-100 Sensor (4-Wire) Calibrate through NABL Lab HTC & CC-2478, Vide Certificate No.: HTC/2020/08/13373. Calibrated on 14-Aug-2020, Traceable to National Standard.
  - 3 Pt-100 Sensor (4-Wire) Calibrate through NABL Lab Nishitronic CC-2294, Vide Certificate No. NI/2008/029/002 Calibrated on 18-Aug-2020. Traceable to National standard.
  - 4 6 1/2 Digital Precision Multimeter Calibrated through NABL Lab HTC & CC-2478, Vide, Certificate No. HTC/2021/04/6227. Calibrated on 29-April-2022, Traceable to National Standard.

23/04/2022  
Shariat S.

The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor  $k=2$ , which corresponds to a coverage probability of approximately 95% for normal distribution.

**Note :**

- 1) UUC stands for Unit Under Calibration.
- 2) This certificate refers only to the particular item submitted for calibration
- 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Hi - Tech Calibration, Vapi".
- 4) The calibration results relate only to the item calibrated reported in the certificate are valid at the time of and under the stated conditions of measurement.

Ankit C Patel  
Calibration Engineer

Calibrated By



Dharmesh R. Purohit  
Quality Manager

Authorised Signatory

HF-31/4



Certificate No. :-	HTC/2021/07/11363	Date of Calibration :-	22-Jul-2021	Page
I.D. No. :-	PVT/SD/TST/31	Recom. Due Date :-	21-Jul-2022	3 of 3
ULR No. :-	CC247821000011363F	Discipline :-	Thermal - Temperature	

## 5. Calibration Method

Actual temperature reading indicated by UUC is compared with specified stable temperature for a given temperature measured using standard – pt-100 Sensor or Thermocouple & DMM or calibrator.

## 6. Calibration Results :

### A) INSTRUMENTAL ERROR FOR TEMPERATURE

Sr. No.	Cal. Point in °C	UUC Reading in °C	Standard Reading in °C	Error in °C	+/- Expanded Uncertainty in °C
1	20.0	20.0	20.075	-0.075	0.820
2	260.1	260.1	260.213	-0.113	0.660
3	750.2	750.2	750.435	-0.235	2.770
4	1000.3	1000.3	1000.588	-0.288	2.530

Ankit C Patel  
Calibration Engineer

Calibrated By  
*AcReal*

Kavita M. Panwala  
Verification Engineer

Checked By  
*KMP*



Dharmesh R. Purohit  
Quality Manager

Authorised Signatory  
*DA*

23/02/2022  
sharad.s.

HF-31C/00



## CALIBRATION CERTIFICATE

Service Request No. :- 2021/07/285		Certificate No. :- HTC/2021/07/11364	
ULR No. :- CC247821000011364F		Certificate Date of Issue :- 28-Jul-2021	
Date of Calibration 22-Jul-2021	Recom.Due Date : 21-Jul-2022	Discipline Thermal - Temperature	NABL Certificate Due on 05-Dec-2021
Page 1 of 3			
1. Customer :- Plasma & Vacuum Technologies Plot No. 17, Road No. 1A Kathwada GIDC, Ahmedabad-382430			
Received Date :- 21-07-2021		Work Instruction No.: HTC/WI/10	
Location of Calibration :- At Lab		Reference Standard : NABL 129 & Euromat cq-8	
Condition of Item :- Good		Environment Condition Temp. °C 24.6 RH % 56.3	
2. Description of Item			
Name :- Thermocouple	Range :- 0 to 1200 °C		
ID No. :- PVT/SD/TST/32	L.C. :- ---		
Sr. No. :- ---	Accuracy :- ---		
Make :- TST	Working Range :- Full		
Model / Type :- --- / K - Type	Location :- PVT/SD/DLD/03		
3. Detail of Master equipment used for calibration			
Name	Make/I.D No.	Certificate No.	Certified By
R-type Thermocouple	HTC-EQP-018	NI/2008/029/001	Nishitronic CC-2294
Pt-100 Sensor (4-Wire)	HTC-EQP-090	HTC/2020/08/13373	HTC, & CC-2478
Pt-100 Sensor (4-Wire)	HTC-EQP-091	NI/2008/029/002	Nishitronic CC-2294
6 1/2 Digital Precision Multimeter	HTC-EQP-017	HTC/2021/04/6227	HTC & CC-2478
Cal. Validity			
17-Aug-2022			
13-Aug-2021			
17-Aug-2022			
28-Apr-2022			
<p>The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor <math>k=2</math>, which corresponds to a coverage probability of approximately 95% for normal distribution.</p> <p>Note :</p> <ol style="list-style-type: none"> <li>1) UUC stands for Unit Under Calibration.</li> <li>2) This certificate refers only to the particular item submitted for calibration</li> <li>3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Hi - Tech Calibration,Vapi".</li> <li>4) The calibration results relate only to the item calibrated reported in the certificate are valid at the time of and under the stated conditions of measurement.</li> </ol>			
<p>Ankit C Patel Calibration Engineer</p> <p><i>ACP</i> Calibrated By</p>		<p>Dhamesh R. Purohit Quality Manager</p> <p><i>DP</i> Authorised Signatory</p>	



23/02/2022  
Shailesh S.

HF-31/4



Name	Make/I.D No.	Certificate No.	Certified By	Page 2 of 3
------	--------------	-----------------	--------------	-------------

All Calibration done in SI units and are traceable to National / International standards as per required ISO/IEC/17025

4. Tracibility :
- 1 R-Type Thermocouple Calibrate through NABL Lab Nishitronic CC-2294, Vide Certificate No.NI/2008/029/001. Calibrated on 18-Aug-2020 Traceable to National Standard.
  - 2 Pt-100 Sensor (4-Wire) Calibrate through NABL Lab HTC & CC-2478, Vide Certificate No.: HTC/2020/08/13373. Calibrated on 14-Aug-2020, Traceable to National Standard.
  - 3 Pt-100 Sensor (4-Wire) Calibrate through NABL Lab Nishitronic CC-2294, Vide Certificate No.NI/2008/029/002 Calibrated on 18-Aug-2020, Traceable to National standard.
  - 4 6 1/2 Digital Precision Multimeter Calibrated through NABL Lab HTC & CC-2478, Vide, Certificate No. HTC/2021/04/6227. Calibrated on 28-April-2022, Traceable to National Standard.

(R) 23/08/2022  
Shaili S.

The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor  $k=2$ , which corresponds to a coverage probability of approximately 95% for normal distribution.

**Note :**

- 1) UUC stands for Unit Under Calibration.
- 2) This certificate refers only to the particular item submitted for calibration
- 3) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Hi - Tech Calibration, Vapi".
- 4) The calibration results relate only to the item calibrated reported in the certificate are valid at the time of and under the stated conditions of measurement.

Ankit C Patel  
Calibration Engineer

Calibrated By



Dharmesh R. Purohit  
Quality Manager

Authorised Signatory

HF-31/4



Certificate No. :-	HTC/2021/07/11364	Date of Calibration :-	22-Jul-2021	Page
I.D. No. :-	PVT/SD/TST/32	Recom. Due Date :-	21-Jul-2022	3 of 3
ULR No. :-	CC247821000011364F	Discipline :-	Thermal - Temperature	

## 5. Calibration Method

Actual temperature reading indicated by UUC is compared with specified stable temperature for a given temperature measured using standard – pt-100 Sensor or Thermocouple & DMM or calibrator.

## 6. Calibration Results :

### A) INSTRUMENTAL ERROR FOR TEMPERATURE

Sr. No.	Cal. Point in °C	UUC Reading in °C	Standard Reading in °C	Error in °C	+/- Expanded Uncertainty in °C
1	20.1	20.1	20.223	-0.123	0.820
2	260.3	260.3	260.445	-0.145	0.650
3	750.4	750.4	750.673	-0.273	2.770
4	1000.5	1000.5	1000.823	-0.323	2.530

Ankit C Patel  
Calibration Engineer

*Ac Patel*  
Calibrated By

Kavita M. Panwala  
Verification Engineer

*KMP*  
Checked By



Dharmesh R. Purohit  
Quality Manager

*Dharmesh R. Purohit*  
Authorised Signatory

*23/07/2022*  
*Shant*

HF-31C/00